

Модули спектрометра UV-NIR, MMS, CGS, MCS FLEX, PGS

Технические характеристики

По вопросам продаж и поддержки обращайтесь:

Алматы (727)345-47-04
Ангарск (3955)60-70-56
Архангельск (8182)63-90-72
Астрахань (8512)99-46-04
Барнаул (3852)73-04-60
Белгород (4722)40-23-64
Благовещенск (4162)22-76-07
Брянск (4832)59-03-52
Владивосток (423)249-28-31
Владикавказ (8672)28-90-48
Владимир (4922)49-43-18
Волгоград (844)278-03-48
Вологда (8172)26-41-59
Воронеж (473)204-51-73
Екатеринбург (343)384-55-89

Иваново (4932)77-34-06
Ижевск (3412)26-03-58
Иркутск (395)279-98-46
Казань (843)206-01-48
Калининград (4012)72-03-81
Калуга (4842)92-23-67
Кемерово (3842)65-04-62
Киров (8332)68-02-04
Коломна (4966)23-41-49
Кострома (4942)77-07-48
Краснодар (861)203-40-90
Красноярск (391)204-63-61
Курск (4712)77-13-04
Курган (3522)50-90-47
Липецк (4742)52-20-81

Магнитогорск (3519)55-03-13
Москва (495)268-04-70
Мурманск (8152)59-64-93
Набережные Челны (8552)20-53-41
Нижний Новгород (831)429-08-12
Новокузнецк (3843)20-46-81
Новосибирск (383)227-86-73
Ноябрьск (3496)41-32-12
Омск (3812)21-46-40
Орел (4862)44-53-42
Оренбург (3532)37-68-04
Пенза (8412)22-31-16
Петрозаводск (8142)55-98-37
Псков (8112)59-10-37
Пермь (342)205-81-47

Ростов-на-Дону (863)308-18-15
Рязань (4912)46-61-64
Самара (846)206-03-16
Санкт-Петербург (812)309-46-40
Саратов (845)249-38-78
Севастополь (8692)22-31-93
Саранск (8342)22-96-24
Симферополь (3652)67-13-56
Смоленск (4812)29-41-54
Сочи (862)225-72-31
Ставрополь (8652)20-65-13
Сургут (3462)77-98-35
Сыктывкар (8212)25-95-17
Тамбов (4752)50-40-97
Тверь (4822)63-31-35

Тольятти (8482)63-91-07
Томск (3822)98-41-53
Тула (4872)33-79-87
Тюмень (3452)66-21-18
Ульяновск (8422)24-23-59
Улан-Удэ (3012)59-97-51
Уфа (347)229-48-12
Хабаровск (4212)92-98-04
Чебоксары (8352)28-53-07
Челябинск (351)202-03-61
Череповец (8202)49-02-64
Чита (3022)38-34-83
Якутск (4112)23-90-97
Ярославль (4852)69-52-93

Россия +7(495)268-04-70

Казахстан +7(727)345-47-04

Беларусь +(375)257-127-884

Узбекистан +998(71)205-18-59

Киргизия +996(312)96-26-47

эл.почта: zsf@nt-rt.ru || сайт: <https://zeiss.nt-rt.ru/>

Introduction

Wavelength ranges

MMS Family
CGS Family
MCS FLEX Family
PGS Family

Software

Areas of application

Definitions and explanations

ZEISS Spectrometer Modules

Compendium of products, electronic components
and software solutions



Information on interactivity on
this page



Page navigation
Previous page | Next page

Introduction

Wavelength ranges

MMS Family
CGS Family
MCS FLEX Family
PGS Family

Software

Areas of application

Definitions and explanations

The moment you discover that your expectations
have been exceeded.

This is the moment we work for.

// SPECTROMETER MODULES
MADE BY ZEISS



Introduction

Wavelength ranges

MMS Family
CGS Family
MCS FLEX Family
PGS Family

Software

Areas of application

Definitions and explanations



Introduction


Your application is our motivation

A traditional spectrometer and/or a traditional monochromator consists of a dispersive medium, an entrance and an exit slit and imaging elements which generate a parallel beam path. To capture a spectrum, a detector behind the exit slit must capture the light sequentially while the dispersive element or the exit slit is moved. This sort of mechanical movement requires time and is prone to interference. However, short measuring times and insensitivity to external influences are quite advantageous for many applications – especially in industry. That is why ZEISS began developing the diode array spectrometers at the end of the 1970s. In place of the exit slit, these spectrometers have a diode array and, through this replacement, capture a complete spectrum simultaneously in a fraction of a second, making moving components unnecessary. The design of the spectrometer module family from ZEISS is based on reducing the optical-mechanical design and the number of components to the physical minimum while using the greatest possible number of identical components for different versions. In the last few years, ZEISS has developed a large number of diverse spectrometer modules for very different applications and requirements. All of these modules offer a key benefit: all spectrometer parts are permanently affixed to each other. This

ensures a very high degree of insensitivity to mechanical vibrations and thus a high level of reliability. Moreover, the entire design is maintenance-free, i.e. recalibration is not necessary. The foundation for the high quality of the spectrometer is the technological know-how at ZEISS for mathematical designs, structuring (grating manufacture and replication), coatings and material processing. Ultimately the joining technology is decisive for ensuring a high degree of insensitivity to influences such as vibrations and, especially, temperature fluctuations.

The following spectrometer module families have been developed at ZEISS:

- **MMS** Monolithic Miniature Spectrometer
- **CGS** Compact Grating Spectrometer
- **MCS FLEX** Multi-Channel Spectrometer
- **PGS** Plane Grating Spectrometer





Click on the respective wavelength range to reach the technical data of the corresponding product.

Introduction


Wavelength ranges

MMS Family
CGS Family
MCS FLEX Family
PGS Family

Software


Areas of application

Definitions and explanations





- Introduction
- Wavelength ranges
- MMS Family
 - Technical Data
 - On-site electronics
- CGS Family
- MCS FLEX Family
- PGS Family
- Software
- Areas of application
- Definitions and explanations



MMS Family

Monolithically compact

The extremely compact design is significant for the spectrometers in the MMS family. Small sizes are available because high repeatability – rather than a high resolving power – is necessary for many applications.

Optical components in the MMS family

- Imaging grating
- A fiber cross-section converter as an optical entrance
- Diode array as an opto-electronic output port

These elements are arranged around and attached to a central body. Depending on the version, the central body is designed as either a glass body or a titanium hollow body. The two components important for the interfaces – the cross-section converter and the detector – are retained.

Central body

On the MMS 1, the central body is a glass body resembling a lens. The imaging grating is replicated directly on this glass body so that it cannot be moved and is optimally protected against dust and gases. An optically denser medium also enables the use of smaller gratings because of the larger aperture, reducing aberrations.

On the UV-sensitive modules, the large glass body has been replaced by a hollow body for reasons of transmission. The grating and detector are affixed to this hollow body. The overall stability is not impaired by the tube design; the temperature-dependent drift of the wavelength has even been reduced.

Gratings

The gratings for the MMS family are holographically blazed flat-field gratings for optimized effectiveness. At ZEISS, these grates are manufactured using the threshold value method and achieve significantly higher effectiveness (for unpolarized light) than sinusoidal gratings. In addition to the dispersive function, the grating must image the entrance slit on the detector array. Via the groove density and curved grooves, coma errors are corrected and the focal curve is evened out (flat field) so that it is optimally adjusted for the flat detector structure. Spectra of over 6 mm long are achieved – even with the small focal length available. Thus the same grating design can be used for the VIS- and the UV-VIS versions. The original grating has an efficiency maximum of approx. 220 nm. The efficiency curve is offset by the factor of the refractive index on the VIS module due to the greater optical thickness.

Cross-section converter

A fiber bundle cross-section converter further optimizes the light intensity. The linear arrangement of individual fibers creates the entrance slit (slit height h determined by the number of individual fibers; the slit width w determined by the core diameter). This is adjusted to the pixel size of the diode array used and to the dispersion properties of the flat-field grating, enabling light intensities to reach the theoretical limit. The cross-section converter is an integral part of the spectrometer design and therefore cannot simply be changed. There is,

however, the possibility of changing the length of the fiber and the design of the entrance. It must also be noted that quartz fibers, such as those used on older MMS UV modules (VIS), create so-called solarization centers when irradiated with deep UV light under 220 nm. This means: the transmission of the fibers is reduced when irradiated with high-energy light. This effect occurs more strongly and more often the shorter the wavelength (higher photon energy), the shorter the intensity and the longer the brightness time. The transmission can also be limited above 220 nm up to 250 nm. This solarization effect can only be partially reversed but can be corrected via frequent reference measurements. For measurements below 225 nm, it is possible to equip the MMS module with solarization-stable fibers. Using a WG 225 filter with 3 mm thickness is an absolute must with standard modules.

Detector

MMS

In the MMS family, the silicon diode array S3904-256Q from Hamamatsu is integrated. Only the MMS 1 NIR enhanced uses the Hamamatsu type S8381-256Q. By using a shorter special housing, the split-off angle is very small, enabling an efficient grating design. This and the 6 mm spectrum length must be considered when switching to another detector. The diode array is coated directly with a dielectric edge filter to suppress the second order.

The following modules are available:

Module	Spectral range (nm)
MMS 1	310 – 1100
MMS UV-VIS	190 – 720 or 250 – 780
MMS UV	195 – 390



MMS 1
Technical Data

Introduction

Wavelength ranges

MMS Family

Technical Data

On-site electronics

CGS Family

MCS FLEX Family


PGS Family

Software

Areas of application

Definitions and explanations


Optical entrance	Input: round	Fiber bundle consists of 30 quartz glass fibers with a 70 µm core diameter, designed as a cross-section converter Diameter: 0.5 mm NA = 0.22 (homogeneous illumination of the acceptance angle) Mounted in an SMA connector
	Output: linear	70 µm x 2500 µm (entrance slit)
Grating	Flat-field, 366 l/mm (in the center)	
Diode array	Manufacturer: Hamamatsu Type: S 3904-256Q in special housing (S 5713) (S 8381-256Q for the MMS 1 NIR enhanced) Number of pixels: 256	
Spectral range	310 nm – 1100 nm Specifications for the range 360 nm – 900 nm (UV-VIS enhanced) 400 nm – 1000 nm (NIR enhanced)	
Wavelength accuracy	0.3 nm	
Temperature drift	< 0.01 nm/K	
Spectral pixel distance	$\Delta\lambda_{\text{pixel}} \approx 3.3$ nm	
Resolution	$\Delta\lambda_{\text{FWHM}} \approx 10$ nm	
Sensitivity	$\approx 10^3$ Vs/J	
Stray light	≤ 0.8 % with halogen lamp for UV-VIS enhanced Transmission at 450 nm with GG 495 filter ≤ 0.2 % with halogen lamp for NIR enhanced Transmission at 650 nm with RG 695 filter	
Dimensions	With housing Cross-section converter (outer length)	70 x 50 x 40 mm ³ Standard: 240 mm, available up to 1 m



Order number	Name	Wavelength range	Description
224001-9001-000	MMS 1 UV/VIS enh.	310 – 1100 nm	PDA with 256 pixels, 240 mm external fiber length
224001-9011-000	MMS 1 UV/VIS enh.	310 – 1100 nm	PDA with 256 pixels, 180 mm external fiber length
000000-1233-038	MMS 1 NIR enh.	310 – 1100 nm	S8381 PDA with 256 pixels, 240 mm external fiber length

MMS UV-VIS I / UV-VIS II
Technical Data

Optical entrance	Input: round	Fiber bundle consists of 30 quartz glass fibers with a 70 µm core diameter, designed as a cross-section converter Diameter: 0.5 mm NA = 0.22 (homogeneous illumination of the acceptance angle) Mounted in an SMA connector
	Output: linear	70 µm x 2500 µm (entrance slit)
Grating	Flat-field, 366 l/mm (in the center), blazed for approx. 220 nm	
Diode array	Manufacturer: Hamamatsu Type: S 3904-256Q in special housing Number of pixels: 256	
Spectral range	UV-VIS I	190 nm – 720 nm Specifications for the 220 nm – 720 nm range
	UV-VIS II	250 nm – 780 nm Specifications for the 250 nm – 780 nm range
Wavelength accuracy	0.5 nm	
Temperature drift	≤ 0.006 nm/K	
Spectral pixel distance	$\Delta\lambda_{\text{pixel}} \approx 2.2$ nm	
Resolution	$\Delta\lambda_{\text{FWHM}} \approx 7$ nm	
Sensitivity	$\approx 10^3$ Vs/J	
Stray light	≤ 0.3 % with deuterium lamp Transmission at 310 nm with NaNO ₂ solution (50 g/l)	
Dimensions	With housing Cross-section converter (outer length)	67 x 60 x 40 mm ³ Standard: 240 mm, available up to 1 m



Order number	Name	Wavelength range	Description
224000-9001-000	MMS UV-VIS I	190 – 720 nm	PDA with 256 pixels, 240 mm external fiber length
000000-1410-176	MMS UV-VIS I	190 – 720 nm	PDA with 256 pixels, 240 mm external fiber length, low solarization
000000-1090-197	MMS UV-VIS II	250 – 785 nm	PDA with 256 pixels, 240 mm external fiber length

MMS UV
Technical Data

Introduction

Wavelength ranges

MMS Family

 Technical Data

 On-site electronics

CGS Family

MCS FLEX Family


PGS Family

Software

Areas of application

Definitions and explanations

Optical entrance	Input: round	Fiber bundle consists of approx. 15 quartz glass fibers with a 70 µm core diameter, designed as a cross-section converter Diameter: 0.4 mm NA = 0.22 (homogeneous illumination of the acceptance angle) Mounted in an SMA connector
	Output: linear	70 µm x 1250 µm (entrance slit)
Grating		Flat-field, 1084 l/mm (in the center), blazed for approx. 220 nm
Diode array		Manufacturer: Hamamatsu Type: S 3904-256N in special housing Number of pixels: 256
Spectral range		195 nm – 390 nm Specifications for the 220 nm – 390 nm range
Wavelength accuracy		0.2 nm
Temperature drift		< 0.005 nm / K
Spectral pixel distance		$\Delta\lambda_{\text{pixel}} \approx 0.8 \text{ nm}$
Resolution		$\Delta\lambda_{\text{FWHM}} \approx 3 \text{ nm}$
Sensitivity		$\approx 10^3 \text{ Vs/J}$
Stray light		$\leq 0.3 \text{ \%}$ deuterium lamp Transmission at 240 nm with NaI solution (10 g/l)
Dimensions	With housing	70 x 60 x 40 mm ³
	Cross-section converter (outer length)	Standard: 240 mm, available up to 1 m



Order number	Name	Wavelength range	Description
224002-9020-000	MMS UV	195 – 390 nm	PDA with 256 pixels, 240 mm external fiber length
000000-1392-178	MMS UV	195 – 390 nm	PDA with 256 pixels, 240 mm external fiber length, low solarization



Introduction

Wavelength ranges

MMS Family

- Technical Data
- On-site electronics
- CGS Family
- MCS FLEX Family
- PGS Family

Software

Areas of application

Definitions and explanations




MMS
On-site electronics

Configuration: an overview



USB / ethernet configuration

USB and ethernet electronics are available for the standard PC interfaces. The USB-based electronics are powered externally through an additional power supply (a self-powered USB device). The PC is connected via a standard USB cable. We recommend a hi-speed USB port (USB 2.0 or 3.0). All electronic circuit boards are designed to be integrated into a customer's housing. The user must provide external ± 12 VDC and + 5 VDC supply voltages.



Introduction

Wavelength ranges

MMS Family

CGS Family

Technical Data

On-site electronics

MCS FLEX Family

PGS Family

Software

Areas of application

Definitions and explanations



CGS UV-NIR Family

More than you'd think

The CGS UV-NIR spectrometers are a class unto themselves. They are extremely compact and robust and are available with a PDA or CCD detector upon request. These spectrometers enable users to measure with maximum quality and optimal spectral efficiency.

Optical components in the CGS family

- Imaging grating
- Optical entrance
- CCD or PDA as an opto-electronic exit port

The CGS comprises an imaging grating, an optical entrance and an uncooled CCD receiver array or a silicon photodiode array (PDA). The CCD receiver array has an electric shutter function which requires minimal integration times and consequently enables high sensitivity. The PDA requires an extremely low noise, ensuring a high signal-to-noise ratio – even in low lighting conditions. The core of the spectrometers is a blazed flat-field grating for light dispersion and imaging. The overall configuration results in a spectral pixel distance of 0.4 nm/pixel with a CCD detector and 0.7 nm / pixel with the PDA detector. A spectral resolution smaller than 3 nm is achieved in accordance with the Rayleigh criterion. The optical entrance is an optical slit on the module side (available in different widths) and an SMA connector on the customer side. All optical components are mounted in a housing made of aluminium.

The spectrometer modules are compact and thermally stable, making them ideal for industrial applications. Their excellent thermal stability and a very low amount of stray light ensure reliable measuring results – even in rough environments. The CGS spectrometer modules extend the MMS and MCS spectrometer module product families.

The new CGS spectrometer combines the benefits of the MMS and MCS spectrometers:

- High resolution
- High sensitivity
- Very good signal-to-noise ratio
- High dynamic range
- Small size

Areas of application

The areas of application for these spectrometers are diverse because of their flexible design. They can be classified in accordance with measurement principles, areas of application or the materials to be analyzed.

Yet their most important advantage is their compactness and insensitivity to external influences so that the modules can be installed in very close proximity to production. An option for on-line inspection is available for most of the applications mentioned below.

The following modules are available:

Module	Spectral range (nm)
CGS UV-NIR CCD	190 – 1015
CGS UV-NIR PDA	190 – 935



Introduction

Wavelength ranges

MMS Family

CGS Family

Technical Data

On-site electronics

MCS FLEX Family

PGS Family

Software


Areas of application

Definitions and explanations

CGS UV-NIR CCD

Technical Data

Optical entrance	SMA connector 50 µm optical slit (can be varied upon request) NA = 0.22 (homogeneous illumination of the acceptance angle) 600 µm mono-fiber interface recommended for customer
Grating	Flat field 534 l/mm (in the center), blazed for approx. 230 nm
Spectral range	190 nm – 1015 nm
Resolution (FWHM) with 50 µm slit	UV-VIS < 2.2 nm NIR < 2.5 nm
Stray light (ASTM 387-04)	3 AU at 240 nm with deuterium lamp (absorption A ₁₀ of NaI)
Integration time (dependent on on-site electronics)	min. 30 µs
Sensor	Hamamatsu S11156, back-thinned CCD, 2048 pixels Detector height: 1 mm Pixel pitch: 14 µm
Housing size L x W x H	78 x 30 x 75 mm ³




Order number	Name	Wavelength range	Description
000000-1794-791	CGS UV-NIR CCD	190 – 1015 nm	Back-thinned CCD, 2048 pixels

CGS UV-NIR PDA

Technical Data

Optical entrance	SMA connector 40 µm optical slit (can be varied upon request) NA = 0.22 (homogeneous illumination of the acceptance angle) 600 µm mono-fiber interface recommended for customer
Grating	Flat field 534 l/mm (in the center), blazed for approx. 230 nm
Spectral range	190 nm – 935 nm
Resolution (FWHM) with 50 µm slit	UV-VIS < 2.0 nm NIR < 2.0 nm
Stray light (ASTM 387-04)	3 AU at 240 nm with deuterium lamp (absorption A ₁₀ of NaI)
Integration time (dependent on on-site electronics)	min. 500 µs
Sensor	Hamamatsu S3903, 1024 pixels
Housing size L x W x H	78 x 30 x 75 mm ³



Order number	Name	Wavelength range	Description
000000-2034-897	CGS UV-NIR PDA	190 – 935 nm	Hamamatsu S3903, NMOS linear image sensor, 1024 pixels


- Introduction
- Wavelength ranges
- MMS Family
 - CGS Family
 - Technical Data
 - On-site electronics
 - MCS FLEX Family
 - PGS Family
- Software
- Areas of application
- Definitions and explanations



CGS UV-NIR PDA

On-site electronics


Configuration: an overview



USB / ethernet configuration

USB and ethernet electronics are available for the standard PC interfaces. The interface electronics (USB and/or ethernet) are powered externally via a power supply unit (self powered). USB-based electronics are connected with the PC via a standard USB cable. A hi-speed USB port (USB 2.0 or 3.0) is required for this configuration.

Ethernet-based configurations are connected to networks via a standard ethernet cable (patch cable) or directly to PCs or laptops via a cross-over ethernet cable. All electronic circuit boards are designed to be integrated into a customer's housing. The user must provide the external + 5 VDC supply voltage.




- Introduction
- Wavelength ranges
- MMS Family
 - CGS Family**
 - Technical Data
 - On-site electronics
- MCS FLEX Family
- PGS Family
- Software
- Areas of application
- Definitions and explanations



CGS UV-NIR CCD

On-site electronics


Configuration: an overview



USB / ethernet configuration


USB and ethernet electronics are available for the standard PC interfaces. The interface electronics (USB and/or ethernet) are powered externally via a power supply unit (self powered). USB-based electronics are connected with the PC via a standard USB cable. A hi-speed USB port (USB 2.0 or 3.0) is required for this configuration.

Ethernet-based configurations are connected to networks via a standard ethernet cable (patch cable) or directly to PCs or laptops via a cross-over ethernet cable. All electronic circuit boards are designed to be integrated into a customer's housing. The user must provide the external + 5 VDC supply voltage.





- Introduction
- Wavelength ranges
- MMS Family
- CGS Family
- MCS FLEX Family**
 - Technical Data
 - On-site electronics
- PGS Family
- Software
- Areas of application
- Definitions and explanations



MCS FLEX Family

Unstoppable

The spectrometers in the MCS FLEX family feature a good resolving power in addition to their high repeatability. All optical components are firmly affixed via a central body, ensuring a robust design.

Optical components in the MCS FLEX family

- Imaging, aberration-corrected grating
- Fiber cross-section converter or slit as an optical entrance
- Diode array and/or a cooled back-thinned CCD as the optoelectronic exit port

In the MCS FLEX family, the different design of the central body determines the system's application. The cross-section converter and detector are used in all the different versions.

Central body

The central body of the MCS FLEX spectrometers consists of a special aluminum alloy to ensure thermal stability (expansion coefficient $\alpha \sim 13 \text{ E}^{-6}$). The aberration-corrected grating, the cross-section converter (or the mechanical slit) as an optical port and the detector are connected via the central body, ensuring excellent stability and reliability. The hollow body means the MCS FLEX can be used for the complete spectrum of the UV-NIR.

Gratings

The gratings for the MCS FLEX family are also holographically blazed flat-field gratings for optimized effectiveness. Maximum grating efficiency has been optimized for different wavelength ranges through additional ion beam etching. Even spectra over a length of 25 nm are achieved through the aberration correction of the gratings. The grating surface is dimensioned in such a way that light from the fiber can be imaged with $\text{NA} = 0.22$.

Cross-section converter

A fiber bundle cross-section converter further optimizes the light intensity. The linear arrangement of individual fibers forms the entrance slit (slit height h is determined by the number of individual fibers; the slit width w is determined by the core diameter). The slit is adjusted to the pixel size of the diode array used and to the imaging dispersion properties of the flat-field grating, enabling light intensities to reach the theoretical limit. The cross-section converter is an integral part of the spectrometer design and therefore cannot simply be altered. There is, however, the possibility of changing the length of the fiber and the entrance design. Please note that quartz fibers, such as those used on older MCS FLEX UV modules (VIS), create so-called solarization centers when irradiated with deep UV light under 220 nm. This means that the transmission of the fibers is reduced when irradiated with high-energy light. This effect is stronger and occurs more often, the shorter the wavelength (higher photon energy), the greater the intensity and the longer the exposure time. The transmission can also be limited above 220 nm up to 250 nm. This solarization effect can only be partially reversed but can be corrected via frequent reference measurements. For measurements below 225 nm, it is possible to equip the MCS FLEX modules with solarization stabilized fibers. Using a WG 225 filter with 3 mm thickness is an absolute must with standard modules.

Detector

MCS FLEX PDA

The MCS FLEX PDA modules use the silicon diode array S3904-1024Q installed by Hamamatsu. The diode array is coated directly with dielectric edge filters to suppress the 2nd order.

Module	Spectral range (nm)
MCS FLEX PDA	190 – 1015

MCS FLEX CCD

Back-thinned CCDs S7031-1006Q from Hamamatsu are installed on the MCS FLEX CCD modules. Back-thinned CCDs are distinguished by direct sensitivity to UV light. To reduce the dark current, this detector has an integrated Peltier element which must be controlled externally. On the MCS FLEX CCD, the warmth discharged by the Peltier element reaches the fan-cooled heat sink via a copper block.

Module	Spectral range (nm)
MCS FLEX CCD	190 – 980



Introduction

Wavelength ranges

MMS Family
CGS Family
MCS FLEX Family

Technical Data
On-site electronics
PGS Family

Software


Areas of application

Definitions and explanations

MCS FLEX PDA

Technical Data

Optical entrance	Cross-section converter
Cross-section converter	Diameter: 0.5 mm NA = 0.22 (consistent illumination of the acceptance angle) Mounted in an SMA connector
Grating	Flat field 248 l/mm (in the center), blazed for approx. 250 nm
Diode array	Manufacturer: Hamamatsu Type: S 3904-1024Q Number of pixels: 1024
Spectral range	190 – 1015 nm
Wavelength accuracy	0.5 nm
Temperature drift	≤ 0.009 nm/K
Spectral pixel distance	$\Delta\lambda_{\text{pixel}} \approx 0.8$ nm
Resolution	$\Delta\lambda_{\text{FWHM}} \approx 3\text{--}4$ nm
Stray light	≤ 0.1 % at 340 nm with deuterium lamp (Transmission of NaNO ₂ solution, 50 g/l, 1 cm)
Housing size L x W x H	160.3 x 62 x 122.2 mm




Order number	Name	Wavelength range	Description
000000-1459-276	MCS FLEX UV-NIR	190 – 1015 nm	PDA with 1024 pixels

MCS FLEX CCD

Technical Data

Optical entrance	Cross-section converter
Cross-section converter	Diameter: 0.5 mm NA = 0.22 (consistent illumination of the acceptance angle) Mounted in an SMA connector
Grating	Flat field 248 l/mm (in the center), blazed for approx. 250 nm
Diode array	Manufacturer: Hamamatsu Type: S 7031-1006 Number of pixels: 1044 x 64
Spectral range	190 – 980 nm
Wavelength accuracy	0.5 nm
Temperature drift	≤ 0.009 nm/K
Spectral pixel distance	$\Delta\lambda_{\text{pixel}} \approx 0.8$ nm
Resolution	$\Delta\lambda_{\text{FWHM}} \approx 3\text{--}4$ nm (UV-NIR version)
Stray light	≤ 0.1 % at 340 nm with deuterium lamp (Transmission of NaNO ₂ solution, 50 g/l, 1 cm)
Housing size L x W x H	198 x 68 x 122 mm (long cross-section converter) 160.3 x 68 x 122 mm (short cross-section converter)



Order number	Name	Wavelength range	Description
000000-1423-352	MCS FLEX CCD UV-NIR	190 – 980 nm	With Hamamatsu CCD detector S7031 with 1024 (1044) x 64 pixels, short cross-section converter
000000-1761-535	MCS FLEX CCD UV-NIR	190 – 980 nm	With Hamamatsu CCD detector S7031 with 1024 (1044) x 64 pixels, long cross-section converter

Introduction

Wavelength ranges

MMS Family
CGS Family
MCS FLEX Family

Technical Data
On-site electronics
PGS Family

Software

Areas of application


Definitions and explanations



MCS FLEX PDA

On-site electronics


Configuration: an overview



USB / ethernet configuration

USB and ethernet electronics are available for the standard PC interfaces. The USB-based electronics are powered externally through an additional power supply (a self-powered USB device). The PC is connected via a standard USB cable. We recommend a hi-speed USB 2.0 port (compatible with a standard USB 1.1).

High-speed USB communication is required to use the fast FEE-1M. All electronic circuit boards designed to be integrated into a customer's housing. The user must provide the external + 5 VDC supply voltage.



Introduction

Wavelength ranges

MMS Family

CGS Family

MCS FLEX Family

Technical Data

On-site electronics

PGS Family

Software

Areas of application

Definitions and explanations



MCS FLEX CCD

On-site electronics


Configuration: an overview



USB / ethernet configuration


USB and ethernet electronics are available for the standard PC interfaces. The USB-based electronics are powered externally through an additional power supply (a self-powered USB device). The PC is connected via a standard USB cable. We recommend a hi-speed USB 2.0 port (compatible with a standard USB

1.1). All electronic circuit boards designed to be integrated into a customer's housing. The user must provide the external + 5 VDC supply voltage.





- Introduction
- Wavelength ranges
- MMS Family
- CGS Family
- MCS FLEX Family
- PGS Family**
 - Technical Data
 - On-site electronics
- Software
- Areas of application
- Definitions and explanations



PGS Family

The NIR specialists

The spectrometers in the PGS family are designed to be used in NIR. InGaAs (indium gallium arsenide) is used as a detector material in this wavelength range. The special combination of an aspheric collimator lens and a focusing lens enables the use of optimized plane gratings for NIR while retaining good flat field correction of the spectral imaging. To ensure long-term stability, all optical components are firmly affixed to each other.

Optical components in the PGS family

- Blazed plane grating
- Aspheric lenses
- Mono-fiber with a slit as an optical entrance
- Cooled InGaAs photodiode array as an optoelectronic output

Central body

In the PGS family, a special aluminum alloy (expansion coefficient $\sim 13E - 6$) is used for the central body. This body is the carrier for the blazed grating and the aspheric collimator and focusing lens. The input fiber and the detector are firmly affixed to the central body, guaranteeing excellent stability.

Gratings

The gratings used in the PGS family are mechanically ruled or holographically exposed. The maximum of the efficiency is modified to the special wavelength range in NIR. With the free diameter, the grating surface is dimensioned in such a way that the light from a fiber can be imaged with a NA of up to 0.37.

Input fiber


The light is generally coupled via a mono quartz fiber. These fibers have a diameter of 600 μm and a NA = 0.22. There is a slit at the end of the fiber with a height of 500 μm (NIR 1.7) and/or 250 μm (NIR 2.2). The slit heights are adjusted to the pixel heights in the InGaAs arrays. A cross-section conversion of the light for creating a higher entrance slit, such as on modules with silicon detectors, is not necessary because of the lower detector height of the InGaAs arrays.

Detector

InGaAs detectors are used in the near infrared range. For the PGS NIR modules, arrays with InGaAs are used for the range up to 1.7 μm and modules with extended InGaAs are used for the range up to 2.2 μm . Arrays are also available with an element number of 256 or 512 (only 1.7 μm) pixels. For the extended InGaAs arrays, an order-sorting filter is applied to the array, depending on the wavelength range, to suppress the 2nd diffraction order.

The following modules are available:

Module	Spectral range (nm)
PGS NIR 1.7-256 UC	960 – 1690
PGS NIR 1.7-256	960 – 1690
PGS NIR 1.7-512	960 – 1690
PGS NIR 2.0-256	1340 – 2000
PGS NIR 2.2-256	1000 – 2150



PGS NIR 1.7 -512

Technical Data

Introduction

Wavelength ranges

MMS Family
CGS Family
MCS FLEX Family
PGS Family


Technical Data
On-site electronics

Software

Areas of application

Definitions and explanations

Optical entrance	Fiber consists of Infrasil quartz glass Diameter: 0.6 mm Length 300 mm NA = 0.22 (homogeneous illumination of the acceptance angle) Mounted in an SMA connector
Input: round	
Output: linear	Slit width: 80 µm
Filter	950 nm edge filter
Grating	Plane grating, 484 l/mm, blazed for approx. 1.2 µm
Diode array	Manufacturer: Hamamatsu Type: S9204 Number of pixels: 512
Spectral range	960 – 1690 nm
Wavelength accuracy	± 1 nm
Temperature drift (10–40°C)	< 0.012 nm/K
Spectral pixel distance	$\Delta\lambda_{\text{pixel}} \approx 1.5 \text{ nm}$
Resolution	$\Delta\lambda_{\text{FWHM}} \approx 7 \text{ nm}$
Stray light	≤ 0.1 % as transmission of 10 mm of water at 1405 nm (measured using a halogen lamp)
Weight	approx. 590 g
Operating temperature	0 – 40°C (standard, depending on cooling electronics)
Storage temperature	-40 – +70°C
Minimal bending radius of fiber (for storage and transport)	50 mm
Minimal bending radius in operation (for wavelength accuracy)	100 mm




Order number	Name	Wavelength range	Description
000000-1307-412	PGS NIR 1.7 t1-512	960 – 1690 nm	NIR spectral sensor, Peltier cooled, InGaAs PDA up to 1.7 µm 512 pixels, dispersion: 1.5 nm/pixel, external fiber length: 300 mm

PGS NIR 1.7-256

Technical Data

Optical entrance	Fiber consists of Infrasil quartz glass Diameter: 0.6 mm Length 300 mm NA = 0.22 (consistent illumination of the acceptance angle), mounted in an SMA connector
Input: round	
Output: linear	Slit width: 80 µm
Filter	950 nm edge filter
Grating	Plane grating, 484 l/mm, blazed for approx. 1.2 µm
Diode array	Manufacturer: Hamamatsu Type: S 9203-256 Number of pixels: 256
Spectral range	960 – 1690 nm
Wavelength accuracy	± 1 nm
Temperature drift (10–40°C)	< 0.012 nm/K
Spectral pixel distance	$\Delta\lambda_{\text{pixel}} \approx 3 \text{ nm}$
Resolution	$\Delta\lambda_{\text{FWHM}} \approx 8 \text{ nm}$
Stray light	≤ 0.1 % as transmission of 10 mm of water at 1405 nm (measured using a halogen lamp)
Weight	approx. 590 g
Operating temperature	0 – 40°C (standard, depending on cooling electronics)
Storage temperature	-40 – +70°C
Minimal bending radius of fiber (for storage and transport)	50 mm
Minimal bending radius in operation (for wavelength accuracy)	100 mm



Order number	Name	Wavelength range	Description
000000-1381-397	PGS NIR 1.7 t1-256	960 – 1690 nm	NIR spectral sensor, Peltier cooled, InGaAs PDA up to 1.7 µm 256 pixels, dispersion: 3.0 nm/pixel, external fiber length: 300 mm

PGS NIR 2.0-256

Technical Data

Introduction

Wavelength ranges

MMS Family
CGS Family
MCS FLEX Family
PGS Family


Technical Data
On-site electronics

Software

Areas of application

Definitions and explanations

Optical entrance	Input: round	Fiber consists of Infrasil quartz glass Diameter: 0.6 mm Length 300 mm NA = 0.22 (homogeneous illumination of the acceptance angle) Mounted in an SMA connector
	Output: linear	Slit width: 80 µm
Filter	1350 nm edge filter	
Grating	Plane grating, 484 l/mm, blazed for approx. 1.4 µm	
Diode array	Manufacturer: Hamamatsu Type: G 9206 Number of pixels: 256	
Spectral range	1340 – 2000 nm	
Wavelength accuracy	± 1 nm	
Temperature drift (10–40°C)	< 0.012 nm/K	
Spectral pixel distance	$\Delta\lambda_{\text{Pixel}} \approx 3 \text{ nm}$	
Resolution	$\Delta\lambda_{\text{FWHM}} \approx 8 \text{ nm}$	
Stray light	≤ 0.1 % as transmission of 10 mm of water at 1405 nm (measured using a halogen lamp)	
Weight	approx. 590 g	
Operating temperature	0 – 40°C (standard, depending on cooling electronics)	
Storage temperature	-40 – +70°C	
Minimal bending radius of fiber (for storage and transport)	50 mm	
Minimal bending radius in operation (for wavelength accuracy)	100 mm	




Order number	Name	Wavelength range	Description
000000-1396-757	PGS NIR 2.0 t2	1340 – 2000 nm	NIR spectral sensor, Peltier cooled Extended InGaAs PDA up to 2.2 µm 256 pixels, dispersion: 1.5 nm/pixel, external fiber length: 300 mm

PGS NIR 2.2-256

Technical Data

Optical entrance	Input: round	Fiber consists of Infrasil quartz glass Diameter: 0.6 mm Length 300 mm NA = 0.22 (consistent illumination of the acceptance angle), mounted in an SMA connector
	Output: linear	Slit width: 80 µm
Filter	950 nm edge filter	
Filter for 2nd order on detector	Yes	
Grating	Plane grating, 300 l/mm, blazed for approx. 1.4 µm	
Diode array	Manufacturer: Hamamatsu Type: G 9206 Number of pixels: 256	
Spectral range	1000 – 2150 nm	
Wavelength accuracy	± 1 nm	
Temperature drift (10–40°C)	< 0.012 nm/K	
Spectral pixel distance	$\Delta\lambda_{\text{Pixel}} \approx 5 \text{ nm}$	
Resolution	$\Delta\lambda_{\text{FWHM}} \approx 16 \text{ nm}$	
Stray light	≤ 0.1 % as transmission of 10 mm of water at 1405 nm (measured using a halogen lamp)	
Weight	approx. 590 g	
Operating temperature	0 – 40°C (standard, depending on cooling electronics)	
Storage temperature	-40 – +70°C	
Minimal bending radius of fiber (for storage and transport)	50 mm	
Minimal bending radius in operation (for wavelength accuracy)	100 mm	




Order number	Name	Wavelength range	Description
000000-1332-256	PGS NIR 2.2 t2	1000 – 2150 nm	NIR spectral sensor, Peltier cooled Extended InGaAs PDA up to 2.2 µm (256 pixels, dispersion: 5 nm/pixel, external fiber length: 300 mm)

- Introduction
- Wavelength ranges
- MMS Family
- CGS Family
- MCS FLEX Family
- PGS Family
 - Technical Data
 - On-site electronics
- Software
- Areas of application
- Definitions and explanations


PGS NIR 1.7-256 UC

Technical Data

Optical entrance	Input: round	FSMA 905 Use NIR-lightguides with core diameter $\geq 600\text{ }\mu\text{m}$, NA = 0.22 to 0.37 Slit width: 80 μm
	Output: linear	
Filter	950 nm edge filter	
Grating	Plane grating, 484 l/mm, blazed for approx. 1.2 μm	
Diode array	Manufacturer: Hamamatsu Type: G9211-01SPL Number of pixels: 256	
Spectral range	960 – 1690 nm	
Wavelength accuracy	$\pm 1\text{ nm}$	
Temperature drift (10–40°C)	$< 0.012\text{ nm/K}$	
Spectral pixel distance	$\Delta\lambda_{\text{pixel}} \approx 3\text{ nm}$	
Resolution	$\Delta\lambda_{\text{FWHM}} \approx 8\text{ nm}$	
Stray light	$\leq 0.1\text{ }\%$ as transmission of 10 mm of water at 1405 nm (measured using a halogen lamp)	
Weight	approx. 590 g	
Operating temperature	0 – 40°C (standard, depending on cooling electronics)	
Storage temperature	-40 – +70°C	



Order number	Name	Wavelength range	Description
000000-2109-070	PGS NIR 1.7-256 UC	960 – 1690 nm	NIR spectral sensor, uncooled Extended InGaAs PDA up to 1.7 μm 256 pixels, dispersion: 3 nm/pixel, external fiber length: 300 mm



- Introduction
- Wavelength ranges
- MMS Family
- CGS Family
- MCS FLEX Family
- PGS Family**
 - Technical Data
 - On-site electronics
- Software
- Areas of application
- Definitions and explanations



PGS


On-site electronics

Configuration: an overview



USB / ethernet configuration

USB and ethernet electronics are available for the standard PC interfaces. The USB-based electronics are powered externally through an additional power supply (a self-powered USB device). The PC is connected via a standard USB cable. We recommend a hi-speed USB 2.0 port (compatible with a standard USB 1.1). The fast FEE-1M requires high-speed USB communication. All electronic circuit boards designed to be integrated into a customer's housing.



Software Solutions

Directly in the process

- Introduction
- Wavelength ranges
- MMS Family
- CGS Family
- MCS FLEX Family
- PGS Family
- Software
- Areas of application
- Definitions and explanations




The architecture of the software products for capturing and processing spectral data is based on a modular structure. This ensures that the software meets diverse, customer-specific specifications and enables different hardware configurations to be adjusted flexibly. For the various operating electronics units, device drivers are available for Windows 2000, XP and Vista.

The universal Aspect Plus program package featuring comprehensive functions is available along with the drivers for the PC bus interface. A programming interface for the SDACQ 32 MP function library is also offered to ensure easy integration into customer-specific applications. This interface directly supports

C/C++/Visual Basic and Delphi, and a LabVIEW™ driver for programming in a LabVIEW™ environment. It is possible to program with finished menu structures for data capture by using the SDPROC32 function library for data capture, configuration and entering parameters.



The SDACQ32MP function library directly addresses these device drivers and supplies a hardware-independent collection of functions, enabling the configuration of the on-site electronics and spectral data capture.



Modular software package for spectral analysis

Aspect Plus

General	Benefits
Aspect Plus is the complex and flexible modular spectral analytical software for MS Windows with special options available as add-ons. Spectral measurements with the spectrometers from ZEISS can be performed and evaluated using Aspect Plus. Comprehensive functions – from the measurement all the way to the formatted printout – simplify analytical evaluation.	<ul style="list-style-type: none">Available in multiple languages (English, French, German, Italian, Portuguese, Spanish), other languages to followMore than one spectrometer can be controlled simultaneouslySupports calibrations (chemometric models) created using standard chemometric software such as GRAMS, UNSCRAMBLER® or UCALFilter function eliminating outlying spectraCommunication via OPC for integration into production line inspectionUse of pre-defined products or creation of user-specific products, as requiredCalculation, evaluation and integration into an upstream process environmentControl of results via Digital I/O



Order number	Name	Description
263259-5020-026	Aspect Plus	Windows spectrometer software
000000-1242-401	Aspect Plus driver for PCs and USBs	Aspect Plus driver for Windows 2000 and XP tec5 electronics

Areas of application

Introduction

Wavelength ranges

MMS Family
CGS Family
MCS FLEX Family
PGS Family

Software

Areas of application

Definitions and explanations

The areas of application for these spectrometers are diverse because of their flexible design. They can be classified in accordance with measurement principles, areas of application or the materials to be analyzed. Compactness and insensitivity to external influences are crucial so that modules can be installed in close proximity to production. An on-line control option is provided in most of the applications mentioned below.

Measuring principles:



- 1. Emission
- 2. Diffuse reflection
- 3. Reflection
- 4. Transmission – absorption
- 5. White-light interference

Emission

A part of the light is injected into the spectrometer to determine the spectral emission of a light source. In many cases, the coupling fiber bundle only needs to be brought close to the light source because of the high light sensitivity. An achromatic converging lens can be used for optimization.

Examples

- Monitoring illuminators (aging)
- Determining the wavelength of LEDs or (tunable) lasers
- Luminescence, fluorescence
- Monitoring the solar spectrum, burns, discharges or plasmas
- Determining the temperature T as per Wien's displacement law: e.g.: 3000 K \longleftrightarrow 966 nm



Requirements


The wavelength accuracy is very high for the size of the module, enabling, through a sub-pixel resolution procedure, an exact identification of the wavelength from light sources which emit a line, e.g. LEDs (calibration). The spectrometer modules are not suitable for analyzing emissions which contain many spectrally adjacent lines.

Diffuse reflection

The diffuse reflection (from rough surfaces) provides information on the color of the surface. In addition to the spectrometer, the light source and the placement (angle to surface normal) of the spectral sensor are important. In most cases, a light source with a wide-band emission is used, e.g. a halogen lamp. In this case, it is usually sufficient to bring the cross-section conversion entrance close to the surface to be measured without an additional optic.

Examples

- Color measurements on diverse surfaces (materials)
- Coating condition
- Determining paper quality



Requirements


The spectrometer modules have been specially designed for color measuring technology. Their high repeatability and light intensity at a moderate spectral resolution meet the specifications exactly.

Reflection

Reflection is a special case of diffuse reflection and refers to the directionally reflected light from 'smooth' low-scatter surfaces. A light source is also required in addition to the sensor. Please note that the reflectivity depends strongly on the alpha angle. A simple setup for measurements under 0°C is possible by using a special light guide which both supplies the light and transmits it to the detector.

Examples

- Coatings in general
- Anti-reflective coatings of surfaces using metals or dielectric coatings
- Ellipsometry
- Determining the fat content in meat and sausages
- Determining the moisture content in humidity in grains, food and cellulose
- Identifying plastics for recycling and disposal



Requirements


Many reflection spectra do not exhibit particularly clear structures. Thus absolute wavelength accuracy is often significantly more important than a good spectral resolving power.

White light interference

Interferences are the result of radiating white light on optically transparent interfaces because, for certain wavelengths, the optical path difference is exactly the multiple of the optical layer thicknesses $n \times d$ (λ_1, λ_2 : position of the extrema; distance: one period). If the refractive index n is known, then the geometric layer thickness d can be determined. The fiber interface ensures easy coupling to microscopes or flanging onto coating systems. Inversely: if layer thickness d is known, then the dispersion n (lambda) can be determined.

Example

- Performing layer thickness measurement on photo resists and dielectric layers



E.g. MMS 1, $n = 1.5$
 $d_{max} \approx 25 \text{ }\mu\text{m}$, $d_{min} \leq 0.2 \text{ }\mu\text{m}$

$2 n \times d = \lambda_1 \times \lambda_2 / (\lambda_1 - \lambda_2)$

Requirements


High absolute accuracy of the wavelength is also necessary to accurately determine the thickness. The maximum measurable thickness is coupled with the spectral resolving power (split of two interference maxima), the minimal thickness with the spectral range to be captured (display of at least a half-period). Absolute intensity values must be known to determine even thinner layers (performing an evaluation of less than a half-period).

Transmission

Radiographing material with the thickness d provides information on the spectral dependence of the absorption constant $\alpha(\lambda)$ (I_0 : radiated intensity, $I(d)$: transmitted intensity). Immersion probes connected to a light source and a spectrometer module via fibers are the simplest way to measure the concentration c of liquids. The concentration is related to the absorption constant via the extinction coefficient ϵ . Otherwise, setting up a collimated beam path is recommended. However, it is also possible to work with the cross-section converter entrance in direct contact with the object to be measured.

Examples

- Measuring filters (color filters, interference filters)
- Determining the concentration of liquids
- Determining the sugar and alcohol content in beverages
- Performing quality assurance in the petrochemical industry



Requirements

In many cases, a very high spectral resolution is once again less important than very good wavelength accuracy and high dynamic resolution, such as that offered by the MMS modules.

Introduction

Wavelength ranges

MMS Family
CGS Family
MCS FLEX Family
PGS Family

Software

Areas of application

Definitions and explanations

Definitions and Explanations of Terms

One of the most important criterion when selecting a spectrometer is the spectral range which the spectrometer must cover. It is usually clear what range is required. However, the two other important criteria for a spectrometer – the spectral and the intensity-related (dynamic) resolution – are not usually clearly defined.

Spectral resolution

The following four terms refer to 'spectral' resolution:

- 1. Rayleigh criterion – $\Delta\lambda_{\text{Rayleigh}}$ (DIN standard)
- 2. Line width, mostly half-value width or full width at half maximum – $\Delta\lambda_{\text{FWHM}}$
- 3. Sub-pixel resolution (also called: 'software resolution')
- 4. Pixel dispersion – $\Delta\lambda/\text{pixel}$

A meaningful definition results from the application. A spectrometer is essentially used to perform three different jobs. These tasks may, of course, overlap:

- 1. Splitting two or more lines within a spectrum – analyzing compounds
- 2. Determining the line form – usually determining the width of a line or a band (FWHM or $1/e^2$ -width)
- 3. Measuring a line with respect to peak wavelength and intensity at the maximum – e.g. determining emissions.

Spectral resolving power


The Rayleigh criterion is relevant for splitting spectral lines as per DIN. This shows how large the spectral distance of two lines $\Delta\lambda_{\text{Rayleigh}}$ must be so that each line can be recognized as separate from the other. The spectral width of the individual lines $\Delta\lambda_{\text{line}}$ (see above) must be significantly less than their distance. This is the only significant definition for the spectral resolving power.

2 lines with $I_{\text{max},1} = I_{\text{max},2}$ are split if
 $\Delta I_{\text{decrease}} \geq 19\%$.

Spectral line width

The widening of the line via the spectrometer must be less than the spectral width of the line itself so that the width of a spectral line $\Delta\lambda_{\text{line}}$ can be measured. It is important to know the expansion $\Delta\lambda_{\text{FWHM}}$ created by the spectrometer. This property is related to the Rayleigh criterion.

$\Delta\lambda_{\text{FWHM}} = \lambda_2(I_{\text{max}}/2) - \lambda_1(I_{\text{max}}/2)$
 $\Delta\lambda_{\text{FWHM}} \approx 0.8 \times \Delta\lambda_{\text{Rayleigh}}$



Wavelength accuracy
To determine the absolute spectral position λ – with a certain accuracy $\Delta\lambda_a$ – of an individual line, a spectrometer with at least this absolute wavelength accuracy $\Delta\lambda_a$ is required. This parameter depends on the position accuracy of the readout elements (pixels or slit/detector) and/or the stability of this position (see below) characterized by the repeatability. In contrast, the absolute wavelength accuracy only depends indirectly on the dispersive and focal properties of the spectrometer and is not a 'resolution' in the traditional sense. The stability (or repeatability) of a spectral sensor depends on the mechanical stability and the temperature-determined wavelength drift. The former is completely noncritical for spectrometer modules and the drift is practically negligible.

Dispersion
The specification $\Delta\lambda / \text{pixel}$ ($= \Delta\lambda_{\text{pixel}}$) has nothing to do with spectral resolution. Instead, it is just the linear dispersion of a diode array spectrometer. Pixel dispersion and spectral resolution are linked via the width of the entrance slit and the imaging properties: if the entrance slit is imaged on approx. 3 pixels, triple the pixel dispersion corresponds approximately to $\Delta\lambda_{\text{Rayleigh}}$.

$\Delta\lambda_{\text{Rayleigh}} \approx 3 \times \Delta\lambda_{\text{pixel}}$



Special features of diode array spectrometers (DAS)

Spectral resolution
Determined by the fixed position of the pixels and/or the wavelength of the radiated light, the resolution is different than on monochromators/spectrometers with movable elements. Resolution – as defined by "splitting two adjacent lines" – depends on the relative position of these lines with respect to the pixels:

If two closely adjacent lines are imaged onto the pixels in such a way that the minimum falls onto the middle pixel (I_3) and the maxima fall on the two neighboring pixels (I_1, I_2), the lines can be split if the displayed intensity is $I_2 < 0.81 \times I_1$ (I_3). $\Delta\lambda$ is exactly two pixels ($2 \times \Delta\lambda_{\text{pixel}}$).

The position of the maxima corresponds relatively exactly to the central wavelengths of the pixels displayed.

If, however, the maximum of a line is imaged onto the dividing line of two pixels (I_1, I_2), then four pixels are required to establish a clear reduction in pixel intensities. Both pixels exhibit roughly the same intensity so that a reduction to 81% is only displayed in the next pixel (I_3). In this case, the real maxima are separated by less than three pixels. However, the DAS displays a spectral distance of $3 \times \Delta\lambda_{\text{pixel}}$ because a diode array only captures discrete values with the step size of the pixel dispersion. A total of four pixels are required for the evaluation.



Sub-pixel resolution
Determining the peak wavelength λ_{max} (and/or peak intensity I_{max}) requires that the spectral line to be measured be imaged onto at least three pixels. With three intensity value pairs per pixel $I_{1,2,3}$ and the central wavelength of the corresponding pixels $\lambda_{1,2,3}$, the line can be e.g. relatively easily modified using a parabola. The parabola equation provides the vertex with the information on the peak wavelength and peak intensity. The accuracy of this method depends primarily on the absolute accuracy of the central wavelength. In principle, this wavelength can be determined with almost any degree of accuracy on a diode array spectrometer. If necessary, each pixel can be individually calibrated. However, stability is crucial. Otherwise, the wavelength specification will only remain valid until the next shock or temperature change. No extreme value determination can be per-

formed if the imaging (and the dispersion) of a DAS is selected in such a way that fewer than three pixels are illuminated. This results in a paradox: a seemingly more advantageous situation – a line is very narrow at the exit – leads to significantly greater inaccuracy. If, for example, a line is imaged into only one pixel, the spectral uncertainty is $\Delta\lambda_{\text{pixel}}$.

Parabola equation
 $I(\lambda) = a \times \lambda^2 + b \times \lambda + c$

Coefficients
 $a = (I_3 + I_1 - 2 I_2) / 2 \Delta\lambda^2$
 $b = (I_3 - I_1) / 2 \Delta\lambda - 2a \times \lambda^2$
 $c = I_2 - a \times I_2^2 - b \times I_2$

Maximum at $\lambda_{\text{max}} = -b / 2a$

Determining the half-value width
The parabola fit also provides qualitative information on the half-value width. To perform a parabola fit, $I_{\text{max}}/2$ just needs to be inserted into the parabola equation. The half-value width of a parabola only deviates slightly from the half-value width of a Gaussian fit.

The half-value width displayed by a DAS depends on the relative position of a line to the individual pixels and is a periodic function of this position with a 1 pixel period length. Our specifications are based on 'worst-case' values.

More suitable – but also more complex – are fits with Gaussian and Lorentz curves which better correspond to the real spectral distributions. These also have the benefit that the resulting calculated half-value width is not dependent on the relative position to the pixels.

$\Delta\lambda_{\text{FWHM}} = 2[(b/2a)^2 - (c - I_{\text{max}}/a)/a]^2$

Intensity resolution

The following properties are of interest for measuring intensities:

- Relative:
 - Smallest detectable change
 - Signal stability
 - Detection range or dynamics
 - Linearity

- Absolute:
 - Lowest detectable light quantity or sensitivity

Introduction

Wavelength ranges

MMS Family
CGS Family
MCS FLEX Family
PGS Family

Software

Areas of application

Definitions and explanations

Accuracy

Measurements of minimal changes and stability depend directly on each other and are essentially determined by the noise within the electronics because most spectrometers ensure a stable 'light path.' As with all sizes, it is important how a value – in the truest sense of the word – is determined. For the spectrometer module specifications, e.g. a 10 ms integration time is selected and the standard deviation $\Delta\sigma$ is calculated above 20 captures. This supplies a measure for the accuracy ΔI which can be used to determine an intensity value.

$$\Delta I = I_{\text{Noise}} = \Delta\sigma$$

Dynamics and intensity changes

The dynamic is understood as the relationship between the saturation level I_{sat} and the noise I_{noise} $\longleftrightarrow \Delta\sigma$ and corresponds to the signal-noise ratio S/N . (The usable range is still reduced by the dark current.) The S/N depends not only on the detector but also the digitization which provides the small step width into which a suitable signal can be separated.

$$\text{Dynamic} = S/N = I_{\text{sat}} / I_{\text{noise}}$$

Of course the weakest link in the chain determines the signal-to-noise ratio to be achieved. With a 14 bit converter e.g. – this corresponds to 16384 steps or increments – and a noise of $\Delta\sigma = 1$ count, a signal (fully controlled) can actually be divided into 16384 steps. The slightest measurable change is thus $1/16384$ of the saturation signal. There is an uncertainty of four counts with a noise of four counts, i.e. only $4/16384$ of the saturation signal can be measured as a definitive change and/or the signal can be meaningfully divided into 4069 steps.

At this point it should be noted that a higher dynamic range is only useful if the detector is adjusted so that it is equally high: you should always try to reach a high level of light so that the high sensitivity of the ZEISS spectrometers is beneficial.

$$\text{Dynamic} = \text{Range ADC} / \Delta\sigma$$

Linearity

These statements only apply to an ideal detector linearity and the connected electronics, i.e. if the measured charge is linearly dependent on the irradiated intensity. The admissible deviation must be specified for quantitative information to be obtained. Fortunately, modern semi-conductor detectors exhibit almost perfect linear behavior over many ranges. Before reaching saturation (the extreme case of non-linearity), however, the increase in the electricity generated (information carrier for intensity) is no longer linear to the number of photons hitting the photosensi-


tive material. The linearity range is consequently smaller than the dynamic range.

External influences

As the graphic shows, a change of temperature T does not cause any change in sensitivity. In the range up to 1100 nm, the sensitivity even increases as the temperature rises. At temperatures between - 50 and + 50 °C, the sensitivity changes by less than 1% in the range of 1 to 1.55 μm , even for InGaAs photodiode arrays. Only outside of the specified range is a stronger temperature influence caused by a different coating. (Falling temperatures cause reduced sensitivity on the band edge.)

The photodiode arrays used do not show any deterioration in the signal-to-noise ratio. Only the dark current I_{dark} increases with rising temperature, resulting in a reduction of the dynamic range. This is why detectors – in particular InGaAs diode arrays – are often cooled. With this in mind, it should be noted that the light quantities to be measured are also subject to fluctuations. The instability of the illumination source is often the limiting factor.

$$I_{\text{dark}}(T+7K) = 2I_{\text{dark}}(T)$$



Sensitivity

The 'smallest detectable change' is a relative specification. It is significantly more difficult to specify the smallest detectable quantity of light at all. Or: how many photons are needed so that the detection electronics detect a change? The difficulties stem from determining the light intensity of a light source and the coupling efficiency. These are also dependent on the wavelength: first because all components have wavelength-dependent efficiencies, including the coupling; second because the bandwidth for the sensitivity measurements are of crucial importance. The simplest case is a light source with a very narrow band, as is featured with most lasers. The situation is at least clear if the bandwidth is significantly smaller than the spectrometer bandwidth. The MMS value of over 10^{13} counts/Ws has been measured with a red HeNe laser.

Stray light

Specifying the stray light value only makes sense in conjunction with the measuring instructions. Stray light values for the spectrometer modules are determined with three different light sources to determine the different spectral components in stray light and/or false light: a deuterium lamp for the UV range and a halogen lamp for the VIS-NIR range.

The stray light level is the ratio between the respective measurement with a GG495 or KG3 filter and the maximum useful signal. Thus the stray light given is for the shortwave range, showing that, on the spectrometer modules, the essential stray light proportion comes from the NIR. This is beneficial because these spectrally 'remote' components can be easily filtered out. For the PGS NIR, the stray light value is reduced to 0.1% (measured with a halogen lamp at 1450 nm, RG 850 filter and 10 mm water absorption).


Stray light affects the dynamic range because the full dynamic range is no longer available due to false light. Changes in the causative radiation only break through in relation to the stray light proportion: e.g. if the stray light proportion is at 1 per mil, a 10% change in the effective radiation means a change of 10^{-4} . If the causative radiation is not used, then the proportion can be further reduced via filtering. In the example described, a blockage of 10^3 leads to a total change of 10^{-7} . There are only small limitations to measuring minuscule changes because the noise is usually much stronger. The stray light proportion can be 'calculated out' if the cause of the signal is known.

Optical interface

Interfaces must be defined mechanically and optically. The SMA plug-in connection – such as that used on all models – is a useful mechanical interface in the optic, resulting in a clear interface along with the well-defined etendue of a fiber bundle.

Etendue

The light etendue G is the product of the light entrance surface F and the opening or spatial angle φ of the light bundle where by the calculation index n must still be observed. The first factor corresponds to the fiber (bundle) cross-section. The second factor is the result of the numeric aperture (NA). In the MMS family, the etendue is calculated at $G = 0.157 \text{ mm}^2\text{sr}$.



$$G = F \times \varphi \times n^2$$
$$\varphi = 2\pi \times (1 - \cos \alpha)$$
$$\alpha = \arcsin \text{NA}$$

In order to optimally modify an existing light source (fiber, lamp, imaging system), it is recommended that the corresponding etendue be determined first. The following coupling efficiency can be estimated through the comparison with the MMS etendue. A 4% Fresnel reflection loss (index jump at the glass fiber) must be observed.

Transmission increase

Assuming the beam is round, then an increase in transmission of $\eta_{\text{F, QSW}} / \eta_{\text{F, SSW}}$ is achieved by using a cross-section converter (CSC) as compared to the classical slit. This can be calculated using the ratio of the light transmitted via the QSW to the light transmitted via a rectangular slit.

With the CSC, the transmitted portion through the fill factor is $\eta_{\text{F, QSW}}$. The fill factor is defined as an optically effective surface A_{eff} with respect to the illuminated entire surface A_{opt} . A_{eff} is, in the case of the QSW, the product of the fiber core cross-section with the diameter d_{fiber} and the number of fibers N at the slit, the surface from slit width b and the slit height h . The entire surface is the circular surface with a diameter $d_{\text{slit}} = h$.

$$\eta_{\text{F, CSC}} = N \times d_{\text{fiber}}^2 / d_{\text{slit}}^2$$
$$\eta_{\text{F, slit}} = 4 \times b / (\pi \times d_{\text{slit}})$$
$$\eta_{\text{F, CSC}} / \eta_{\text{F, slit}} = 16 \text{ (MMS)}$$

Diode array spectrometer optimization

In addition to selecting extremely efficient components (a blazed grating, a cross-section converter, a sensitive diode array), the dispersion, imaging properties, entrance slit, pixel size and pixel distances must be well-matched. It is crucial for light sensitivity that – with monochromatic light – more than 2 pixels are illuminated for the spectral resolution. The grating images 1:1 in the first approximation, e.g. the entry slit should be 2 to 3 pixels wide. If more pixels are illuminated, the signal-to-noise ratio and the sensitivity become worse (1 pixel captures a bandwidth that is too narrow). If fewer than 3 pixels are illuminated, the wavelength accuracy becomes worse. That is why e.g. the selection of 70 μm individual fibers for the QSW on the MMS modules are nearly perfect for a pixel width of 25 μm . The number of fibers is the result of the pixel height divided by the external diameter of the individual fibers.

Introduction

Wavelength ranges

MMS Family
CGS Family
MCS FLEX Family
PGS Family

Software

Areas of application

Definitions and explanations

The moment you achieve absolute confidence.
This is the moment we work for.

// SPECTROMETER MODULES
MADE BY ZEISS



Алматы (727)345-47-04
Ангарск (3955)60-70-56
Архангельск (8182)63-90-72
Астрахань (8512)99-46-04
Барнаул (3852)73-04-60
Белгород (4722)40-23-64
Благовещенск (4162)22-76-07
Брянск (4832)59-03-52
Владивосток (423)249-28-31
Владикавказ (8672)28-90-48
Владимир (4922)49-43-18
Волгоград (844)278-03-48
Вологда (8172)26-41-59
Воронеж (473)204-51-73
Екатеринбург (343)384-55-89

Иваново (4932)77-34-06
Ижевск (3412)26-03-58
Иркутск (395)279-98-46
Казань (843)206-01-48
Калининград (4012)72-03-81
Калуга (4842)92-23-67
Кемерово (3842)65-04-62
Киров (8332)68-02-04
Коломна (4966)23-41-49
Кострома (4942)77-07-48
Краснодар (861)203-40-90
Красноярск (391)204-63-61
Курск (4712)77-13-04
Курган (3522)50-90-47
Липецк (4742)52-20-81

Магнитогорск (3519)55-03-13
Москва (495)268-04-70
Мурманск (8152)59-64-93
Набережные Челны (8552)20-53-41
Нижний Новгород (831)429-08-12
Новокузнецк (3843)20-46-81
Ноябрьск (3496)41-32-12
Новосибирск (383)227-86-73
Омск (3812)21-46-40
Орел (4862)44-53-42
Оренбург (3532)37-68-04
Пенза (8412)22-31-16
Петрозаводск (8142)55-98-37
Псков (8112)59-10-37
Пермь (342)205-81-47

Ростов-на-Дону (863)308-18-15
Рязань (4912)46-61-64
Самара (846)206-03-16
Санкт-Петербург (812)309-46-40
Саратов (845)249-38-78
Севастополь (8692)22-31-93
Саранск (8342)22-96-24
Симферополь (3652)67-13-56
Смоленск (4812)29-41-54
Сочи (862)225-72-31
Ставрополь (8652)20-65-13
Сургут (3462)77-98-35
Сыктывкар (8212)25-95-17
Тамбов (4752)50-40-97
Тверь (4822)63-31-35

Тольятти (8482)63-91-07
Томск (3822)98-41-53
Тула (4872)33-79-87
Тюмень (3452)66-21-18
Ульяновск (8422)24-23-59
Улан-Удэ (3012)59-97-51
Уфа (347)229-48-12
Хабаровск (4212)92-98-04
Чебоксары (8352)28-53-07
Челябинск (351)202-03-61
Череповец (8202)49-02-64
Чита (3022)38-34-83
Якутск (4112)23-90-97
Ярославль (4852)69-52-93

Россия +7(495)268-04-70

Казахстан +7(727)345-47-04

Беларусь +(375)257-127-884

Узбекистан +998(71)205-18-59

Киргизия +996(312)96-26-47

эл.почта: zsf@nt-rt.ru || сайт: <https://zeiss.nt-rt.ru/>